

## Technical data Sheet I-Stream GL

Product code: ISP60GL + diopter

Model: I-Stream GL


Manufacturer: Md Tech Srl - Via Fratelli Bandiera, Ncc - 80026 Casoria (Na), Campania Italia

Intended Use: Cataract surgery

Use: Intraoperative ophthalmic use

Italian Registration Number: 496132

One-piece hydrophobic acrylic lens with aspherical front surface with Glare Stop Barrier filter.



### // General Features

<b>Material</b>	Hydrophobic acrylic with UV filter with water content <1%, biocompatible with eye tissue
<b>Lens Geometry</b>	Biconvex optic, Monofocal, Aspheric, Square Edge 360°.
<b>Dioptric Range</b>	from -5.0D to + 10.0D with increment of 1D; from +10.5D to +30.5D with increment of 0.5 Dpt; from +31.0 to 40.0 with increment of 1D
<b>Incision</b>	from 2,2mm with single use injector

### // Optical Features

<b>Optic Diameter</b>	6.0mm
<b>Total Diameter</b>	13.0mm
<b>Sferical Aberration</b>	-0,26 $\mu$ m
<b>Color</b>	Clear / Natural Yellow
<b>Refractive Index</b>	1,56
<b>UV Filter</b>	White UV CUT OFF < 10% @370nm
<b>A Costant</b>	118
<b>Constant with Ulib System</b>	SRK/T = 118,8; SRK/2 = 119,1; Holladay1 sf = 1,66; Holladay 2 = 5,432; Barrett = 1,78 HofferQ pACD = 5,45; Haigis a0 = 1,26; a1 = 0,40; a2 = 0,10

## // Haptic Features

Type	Optical disc and haptics without profile discontinuity
Haptic Angulation	0° / 5°

## // Other Specifications

Reference Standard	Directive 93/42 CEE, UNI EN ISO:11979
Class	II b Certiquality Srl, Notified Body n°0546, certif. N.18539/1/I
Control Performed	DIOPTER , MTF, Cosmetic
Packaging	Each lens is packed in a lens case polypropylene blister closed closed in a tyvek pouch. Each pouch is contained in a carton box with administrative label.
Product Type	Medical device in compliance with the current legislation
Compatibility With Solutions	Product can be used with hyaluronic acid, balanced salt solution. It must not be used with acid or melted with other substances.
Shelf Life And Sterilization	Product has 60 months of validity from sterilization date and until the date declared on label. Sterilized with EtO according to UNI EN ISO 11135. Product cannot be re-sterilized.
Tests	In vitro cytotoxicity tests, allergic sensitization tests on guinea pigs, eye irritation tests on guinea pigs, product stability tests have been carried out on the product in the experimental phase (reports available in the company)
Presence of Latex	No
Storage	Store at ambience condition, in dry place without humidity. Temperature must be in the range +5°C and+45°.
Information Reported on the Packaging	Batch number, expiry date, serial number, sterilization, manufacturer, CE mark.

